PT. INNOVATION HEALTHCARE INDONESIA

LAICA
ADVANCED TECHNOLOGY

TECHNICAL SPECIFICATIONS

Commercial code	EA1003
Model	JH-117
Output. OSPL90 Max.	≤135dB+3dB
FOG50 MAX Gain	≤40dB+3dB
The full-on gain value (HFA FOG@50)	35±5dB
EQ Input Noise	≤35dB
Frequency Range (Not narrower than)	300Hz~3200Hz
Working Current	≤10mA
Total Harmonic Distortion	≤15%
Rated Volatage	d.c 1.5v
Battery Size	A675
Operating temperature/humidity	-10℃~+40℃ (14℉-104℉), ⩾30%RH,⩽75%RH
Storage temperature/humidity	-10℃~+55℃ (14℉-131℉), ⩾10%RH,⩽90%RH

BODY AID HEARING AMPLIFIER

CODE : **EA1003**





PRODUCT PLUS

- Max. sound output 135 dB
- Average gain 35 dB
- Low noise (<35 dB)
- Frequency range 300-3500 Hz
- Suitable for mild hearing loss

SURGERY for CHILDREN

B.L.R.D.

PT. INNOVATION HEALTHCARE INDONESIA
Komplek Ruko Permata Kota, Blok Q No. 1,1Jln. Pangeran Tubagus Angke No. 170 Jakarta Utara 14450
Phone : (021) 26075090, 26080866 Fax 021-26080867 Email : onlinesales1.ihi@gmail.com
Website : www.innovationindo.com | WhatsApp : +62852 1795 0328